

REPORT

~~CP/NO~~

DATE DISTR. 21 June 51

NO. OF PAGES 2

NO. OF ENCLS.
(LISTED BELOW)

SUPPLEMENT TO
REPORT NO.

25X1X

tubes of U.S.A.

CONFIDENTIAL

CONFIDENTIAL

[illegible]

CONFIDENTIAL
CONFIDENTIAL

- 2 -

- b. The X-ray tubes showed reasonable competence on the part of their producers and were of the kind usually produced to meet a price by manufacturers of considerable tube making ability.
- c. The steel jacketing on the anode is considered to represent inferior practice; it may have been used to diffuse heat, thus preventing cracks in the anode copper, or it may have been used in molding the copper and simply not removed before assembling the tubes.
- d. Inspection has revealed no unusual features warranting further analysis.

- end -

CONFIDENTIAL

CONFIDENTIAL